LITHIUM ION BATTERY TEST SUMMARY

UN 38.3 Lithium Battery Test Summary for ergoflix® basic/FlixOne Battery

1	**Battery Manufacturer** iangsu Naige New Energy Technology Co., Ltd. No. 64, Dongfeng road, Wei Miao Zhen, Pei County, Xuzhou City, Jiangsu Province, China
2	**Product Manufacturer** Anhui JBH Medical Apparatus Co., Ltd. No. 116 Qicang Road, Mingguang, Chuzhou Anhui Province, Anhui, 239400 CHINA +86 85 78 08 66 info@jbhmedical.com
3	**UN38.3 Test Lab** ATS Electronic Technology Co., Ltd. 3/F., Building A, No.1 Hedong 3rd Road, Jinxia Community, Changan, Dongguan, Guangdong, China Tel.: + 86 769 3897 5958 www.dgats.com ats@dgats.com
4	**Test Report Number** ATSU230502711
5	**Date of Test Report** 2023-05-20
6	**Description of Cell or Battery** The Li-ion Battery Model NG00212 25.2V, 11400mAh, 287,28Wh L: 330mm, W: 100mm, H: 105mm
7	**UN38.3 Tests Performed and Successfully Passed** T.1, T.2, T.3, T.4, T.5, T.6, T.7, T.8
8	**Assembled Battery Testing Requirements** Not Applicable
9	**Edition of UN Manual of Tests and Criteria Used** UNITED NATIONS „Manual of Tests and Criteria“, Seventh revised edition (ST/SG/AC.10/11/Rev.7), Section 38.3: Lithium metal and Lithium ion batteries
10	**Company stamp** ergoflix **ergoflix Group GmbH** Marktstr. 20 · 46499 Hamminkeln info@ergoflix.de · www.ergoflix.de Tel. 02852 94590-00 · Fax 02852 94590-99